

# Sensor Certification

**SENSIRION**  
THE SENSOR COMPANY

Product: SHTxx Family

Description: Digital Humidity and Temperature Sensor

This is to certify that above product is calibrated using dedicated equipment which is calibrated according to ISO/IEC 17025 and calibrated by comparison against natural standards, natural physical constants, consensus standards, or by ratio type measurements using self-calibrating techniques.

Standards incorporated to establish this comparison are NIST and NPL traceable.

NIST Traceable Accuracy Information: Range: 50%-90% RH, Accuracy: +/- 2% RH



A handwritten signature in black ink, appearing to be "d. G.", positioned above a horizontal line.

CEO, Sensirion AG



A handwritten signature in black ink, appearing to be "P. Finckel", positioned above a horizontal line.

Quality Manager, Sensirion AG



The sensors included with this package were calibrated in June 2009 and will expire in June 2011